

## EU-declaration of Conformity

**Manufacturer:** Meyland-Smith A/S  
Gøteborgvej 8-12  
9200 Aalborg  
Denmark  
Phone: +45 98961985  
CVR: 26921465  
SRN: DK-MF-000002369

**Hereby declare that**

Basic UDI-DI (GMN):	57143390009PY
Product(s):	See table on page 2
Risk Class:	Class I (according to rules set out in Annex VIII)

**Are in conformity with**

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017

**And have been tested under**

EN 11199-3 Walking aids manipulated by both arms – Requirements and test methods – Part 3: Walking tables  
EN/ISO 10993-5 Biological evaluation of medical devices – Part 5: tests for cytotoxicity

**And therefore, satisfies all essential requirements with reference to**

EN12182 Assistive products for persons with disability – General requirements and test methods.

**And a risk analysis has been made according to**

EN ISO 14971 Medical devices – Application of risk management to medical devices.

On behalf of Meyland-smith A/S  
Aalborg, April 28<sup>th</sup>, 2021



Mathias Haslund Jungersen  
Teamleader, Product Engineering

Basic UDI-DI 57143390009PY includes the following products and their associated configurations:

Item no.	Product	GMDN Code	UDI-DI (GTIN)	Max. Load
M011200	Meywalk 4, Medium, rød	38704	5714339902601	100 kg
M011250	Meywalk 4, Large, sort	38704	5714339902618	130 kg
M013000	Meywalk 4, Small, grøn	38704	5714339903851	60 kg
M014000	Meywalk 4, Micro, sort	38704	5714339904087	60 kg
MW104175	Meywalk 2000, Medium	38704	5714339909235	100 kg
MW104173	Meywalk 2000, Large	38704	5714339909228	100 kg
MW104500	Meywalk 2000, Small, grøn	38704	5714339909242	100 kg
MW106061	Miniwalk str. 1	38704	5714339909259	40 kg
MW106062	Miniwalk str. 2	38704	5714339909266	40 kg